

Buccutite™ Rapid PE-iFluor® 594 Tandem Antibody Labeling Kit *Microscale Optimized for Labeling 100 ug Antibody Per Reaction*

Catalog number: 1356
Unit size: 2 Labelings

Component	Storage	Amount (Cat No. 1356)
Component A: Buccutite™ FOL-Activated PE-iFluor® 594	Refrigerated (2-8 °C), Minimize light exposure	2 Vials (Lyophilized)
Component B: Buccutite™ MTA	Minimize light exposure, Refrigerated (2-8 °C)	2 Vials (Lyophilized)
Component C: Reaction Buffer	Minimize light exposure, Refrigerated (2-8 °C)	1 Vial (20 uL)

OVERVIEW

The Buccutite™ Rapid PE-iFluor® 594 Tandem Antibody Labeling Kit provides a streamlined approach for efficient, small-scale labeling of antibodies with PE-iFluor® 594. Compared to conventional methods such as SMCC crosslinking, Buccutite™ technology offers a more straightforward and reproducible solution. Using a simple two-step mixing protocol, antibodies or proteins can be conjugated with PE-iFluor® 594 in less than two hours. Each kit includes all reagents necessary for two labeling reactions, with Buccutite™ FOL-Activated PE-iFluor® 594 vials specifically formulated to label 100 µg of purified protein or antibody per reaction. Prior to labeling, stabilizing proteins (e.g., BSA) should be removed, and amine-rich buffers such as Tris should be avoided to prevent interference with the labeling chemistry.

PE-iFluor® 594 is a tandem fluorophore with excitation and emission maxima at ~565 nm and ~606 nm, respectively. Its high fluorescence intensity makes it particularly suitable for detecting low-abundance targets while minimizing spectral spillover and reducing compensation complexity. These properties make PE-iFluor® 594 an excellent choice for flow cytometry, spectral flow cytometry, and other immunoassays requiring high sensitivity. However, it is not recommended for applications where photostability is critical. This kit enables direct conjugation of primary antibodies, eliminating the need for secondary antibody labeling strategies. The resulting conjugates streamline workflows and facilitate the development of complex multicolor assays, enhancing experimental flexibility and reducing reagent complexity.

AT A GLANCE

Protocol Summary

1. Add 5 µL Reaction Buffer (Component C) into antibody (100 µL)
2. Add the antibody solution into Buccutite™ MTA vial (Component B)
3. Incubate at room temperature for 30 minutes
4. Mix with 50 µL Buccutite™ FOL-Activated PE-iFluor® 594 (Component A)
5. Incubate at room temperature for 60 minutes

Important: Store the kit at 4 °C upon receipt. When stored correctly, the kit remains stable for six months. Alternatively, Component B can be stored at -20 °C. Avoid freezing Buccutite™ FOL-Activated PE-iFluor® 594 (Component A) and Reaction Buffer (Component C). Before use, warm all components to room temperature and briefly centrifuge the vials before opening. Prepare the necessary solutions immediately prior to starting the conjugation. The following SOP provides an example for labeling goat anti-mouse IgG antibody.

PREPARATION OF WORKING SOLUTION

Antibody Working Solution

1. To label 100 µg of antibody (assuming a concentration of 1 mg/mL), mix 5 µL (5% of the total reaction volume) of Reaction Buffer (Component C) with 100 µL of the antibody solution.

Note: If your antibody has a different concentration, adjust the volume to ensure approximately 100 µg of antibody is available for the labeling reaction.

Note: The antibody should be dissolved in 1X phosphate buffered saline (PBS), pH 7.2-7.4; If the antibody is dissolved in glycine buffer, it must be dialyzed against 1X PBS, pH 7.2-7.4, or use ReadUse™ 10KD Spin Filter (Cat. # 60502 from AAT Bioquest) to remove free amines or ammonium salts (such as ammonium sulfate and ammonium acetate) that are widely used for antibody precipitation.

Note: Impure antibodies or antibodies stabilized with bovine serum albumin (BSA) or gelatin will not be labeled well.

Note: The antibody-Buccutite™ MTA reaction efficiency is significantly reduced if the antibody concentration is less than 1 mg/mL. For optimal labeling efficiency the final antibody concentration range of 1-10 mg/mL is recommended.

SAMPLE EXPERIMENTAL PROTOCOL

Run Antibody-Buccutite™ MTA reaction

1. Add the antibody working solution directly into the vial of Buccutite™ MTA (Component B), and mix them well by repeatedly pipetting for a few times or vortex the vial for a few seconds.
2. Keep the antibody-Buccutite™ MTA reaction mixture at room temperature for 30 - 60 minutes.

Note: The antibody-Buccutite™ MTA reaction mixture can be rotated or shaken for longer time if desired.

Make antibody-PE-iFluor® 594 Conjugation

1. Make Buccutite™ FOL-Activated PE-iFluor® 594 solution by adding 50 µL ddH2O into the vial of Buccutite™ FOL-Activated PE-iFluor® 594 (Component A), mix well by repeatedly pipetting for a few times or vortex the vial for a few seconds.
2. Mix whole vial of Buccutite™ FOL-Activated PE-iFluor® 594 solution into the antibody-Buccutite™ MTA solution, mix well and rotating the mixture for 1 hour at room temperature.
3. The antibody-PE-iFluor® 594 conjugate is now ready to use.

Note: For immediate use, the antibody-PE-iFluor® 594 conjugate need be diluted with the buffer of your choice.

Storage of Antibody-PE-iFluor® 594 Conjugate

The antibody conjugate should be stored at > 0.5 mg/mL in the presence of a carrier protein (e.g., 0.1% bovine serum albumin). The Antibody-PE-iFluor® 594 conjugate solution could be stored at 4 °C for two months without significant change when stored in the presence of 2 mM sodium azide and kept from light. For longer storage, the antibody-PE-iFluor® 594 conjugates could be lyophilized and stored at ≤ -20 °C.

Table 1. Available fluorophores at AAT Bioquest Buccutite™ Rapid Antibody Labelling Kits

Cat#	Labels	Ex (nm)	Em (nm)
1325	PerCP	482	677
1310	PE	565	575
1318	PE-Texas Red	565	600
1356	PE-iFluor® 594	565	606
1322	PE-Cy5	565	674
1316	PE-Cy5.5	565	700
1358	PE-iFluor® 710	565	710
1317	PE-Cy7	565	780
1311	APC	651	662
1320	APC-Cy5.5	651	700
1319	APC-iFluor® 700	651	713
1321	APC-Cy7	651	780

EXAMPLE DATA ANALYSIS AND FIGURES




Figure 1. AAT Bioquest offers the Buccutite™ rapid labeling kit to streamline PE tandem dye conjugation for antibodies and other proteins, including streptavidin and secondary reagents. This kit utilizes preactivated PE modified with Buccutite™ FOL, while your antibody or protein is modified with Buccutite™ MTA to produce MTA-modified proteins. The MTA-modified proteins react efficiently with FOL-modified PE, yielding the desired PE-antibody conjugate with significantly higher efficiency compared to traditional SMCC chemistry. Additionally, the reaction requires much lower biopolymer concentrations, enhancing efficiency and reducing material usage compared to SMCC-based methods.

DISCLAIMER

AAT Bioquest provides high-quality reagents and materials for research use only. For proper handling of potentially hazardous chemicals, please consult the Safety Data Sheet (SDS) provided for the product. Chemical analysis and/or reverse engineering of any kit or its components is strictly prohibited without written permission from AAT Bioquest. Please call 408-733-1055 or email info@aatbio.com if you have any questions.